



Field Research and Consultation Group

10/26/17

Tamas Ugrai
University of Washington
Oceanography, Box 355351
Seattle, WA 98195-5351

FRCG #: 2017-54

Dear Tamas,

Enclosed within this document are the results of our 10/25/17 clean room evaluation of the Rm 443A Laboratory located in at the University of Washington in the Oceanography Building in Seattle, WA.

According to ISO 14644, we have concluded that at the time and conditions of the measurements, the lab complies with the ISO Class 6 @ 0.3 μm classification. This is equivalent to the now obsolete Federal Standard 209e Class 1000 @ 0.3 μm . Classification is based on each location passing the designated criteria independently (In contrast to Fed Standard 209e, which assesses the overall cleanliness of the room). Therefore, the overall classification was determined by the location with the highest concentration (Sample 4).

At the time of the evaluation, the laminar flow hood and flow bench were operating. Two individuals were present: a laboratory employee, conducting typical work near the sink and fume hood; and the industrial hygienist responsible for monitoring. Please see the attached report and Table 1 and Figure 1 for details.

For comparison, the results of the 04/02/14 evaluation are included with this report (Table 2). This evaluation found that the laboratory complied with ISO Class 6 @ 0.5 μm (Fed Standard 209e Class 1000 @ 0.5 μm).

Lastly, Table 3 provides a comparison of the ISO 14644 and Federal Standard 209e criteria.

Please feel free to contact me with any questions at (206) 616-7689.

Sincerely,

Marc Beaudreau, MS
Industrial Hygienist

Clean Room Test Report

Date and time of testing: October 26, 2017

| | | | |
|-------------------------|---------------------------------------|-------------------------------|-----------------------|
| Laboratory: | UW Oceanography Building Room 443A | Applicable Standard: | ISO 14644 |
| Instrument: | TSI AeroTrak 9306-V2 | Target Criteria: | ISO 6 (@ 0.3 μ m) |
| Room Dimensions: | 5.5m x 3.0m x 2.7m | Minimum sample volume: | 2 Liters |
| Room Area: | 16.7 m ² | Minimum sample number: | 25 |
| Flow Regime: | Non-unidirectional | Flow rate: | 2.83 LPM |
| State of Room: | Operational | Sample duration: | 2:00 min |
| | | Volume/sample: | 0.01 m ³ |
| | | Number of locations: | 5 |
| | | Replicates/location: | 5 |

Table 1. Particle concentration in Oceanography lab 443A on 10/25/17

| Location* | Particle Concentrations (particles/m ³) | | | | # of Samples |
|--------------------|---|---------|--------------|---------|--------------|
| | >0.3 μ m | | >0.5 μ m | | |
| | Average | Std Dev | Average | Std Dev | |
| Sample 1 | 8,525 | 5,492 | 6,979 | 4,755 | 5 |
| Sample 2 | 4,735 | 3,853 | 3,993 | 3,703 | 5 |
| Sample 3 | 495 | 521 | 283 | 387 | 5 |
| Sample 4 | 19,364 | 18,846 | 6,961 | 5,344 | 5 |
| Sample 5 | 177 | 395 | 71 | 158 | 5 |
| ISO Class 5 | 10,200 | | 3,520 | | |
| ISO Class 6 | 102,000 | | 35,200 | | |
| ISO Class 7 | 1,020,000 | | 352,000 | | |

* Refer to Figure 1 for sample locations

Figure 1. Lab schematic and sample locations




Table 2. Particle concentrations in Oceanography lab 443A on 04/02/14

| Location | Particle Concentrations (particles/m ³) | | | | # of Samples |
|--|---|---------|---------|---------|--------------|
| | >0.3 um | | >0.5 um | | |
| | Average | Std Dev | Average | Std Dev | |
| A (near window) | 27,220 | 14,577 | 20,760 | 10,557 | 3 |
| B | 11,318 | 8,292 | 8,834 | 6,356 | 3 |
| C | 12,478 | 9,277 | 10,380 | 8,040 | 3 |
| D | 11,926 | 5,697 | 9,386 | 4,194 | 3 |
| E (near entrance) | 9,441 | 1,791 | 7,067 | 1,675 | 3 |
| Overall | 14,477 | 10,109 | 11,285 | 7,689 | 15 |
| Criteria for M4.5 (Class 1000 or ISO 6) | | | 35,300 | | |
| Criteria for M4 | 30,900 | | 10,000 | | |
| Criteria for M3.5 (Class 100 or ISO 5) | 10,600 | | 3,530 | | |

Table 3. ISO 14644 comparison to Federal Standard 209e

| Class | Max particles/m ³ | | | | | | Fed STD 209e equivalent |
|-------|------------------------------|--------|--------|----------|---------|--------|-------------------------|
| | ≥0.1 | ≥0.2 | ≥0.3 | ≥0.5 | ≥1 | ≥5 | |
| ISO1 | 10 | 2.37 | | | | | |
| ISO2 | 100 | 23.7 | 10.2 | 3.5 | | | |
| ISO3 | 1000 | 237 | 102 | 35 | 8.3 | | Class 1 |
| ISO4 | 10000 | 2370 | 1020 | 352 | 83 | | Class 10 |
| ISO5 | 100000 | 23700 | 10200 | 3520 | 832 | 29 | Class 100 |
| ISO6 | 1000000 | 237000 | 102000 | 35200 | 8320 | 293 | Class 1000 |
| ISO7 | | | | 352000 | 83200 | 2930 | Class 10000 |
| ISO8 | | | | 3520000 | 832000 | 29300 | Class 100000 |
| ISO9 | | | | 35200000 | 8320000 | 293000 | Room Air |

Appendix A

TSI P/N 2300157



CERTIFICATE OF CALIBRATION

TSI Incorporated, 500 Cardigan Road, Shoreview, MN 55126 USA
Tel: 1-800-874-2811 1-651-490-2811 Fax: 1-651-490-3824 <http://www.tsi.com>

SIZE CALIBRATION AND VERIFICATION OF SIZE SETTING

| NOMINAL PARTICLE SIZE | GAIN STAGE | DIGITAL CUTPOINT | EXPANDED UNCERTAINTY |
|-----------------------|------------|------------------|----------------------|
| 0.3 µm | A | 29 | 4.1% |
| 0.5 µm | A | 354 | 3.9% |
| 1 µm | B | 8 | 3.9% |
| 3 µm | B | 80 | 3.6% |
| 5 µm | B | 179 | 3.6% |
| 10 µm | B | 588 | 3.7% |

COUNTING EFFICIENCY

| PARTICLE SIZE | ACTUAL | ALLOWABLE RANGE | PASS/FAIL |
|---------------|--------|-----------------|-----------|
| 0.3 µm | 51% | 50% ± 20% | Pass |
| 0.5 µm | 103% | 100% ± 10% | Pass |

SIZE RESOLUTION

| PARTICLE SIZE | MEASURED | ALLOWABLE RANGE | PASS/FAIL |
|---------------|----------|-----------------|-----------|
| 0.5 µm | 5.6% | ≤15% | Pass |

FALSE COUNT RATE

| SAMPLE TIME (MIN) | SAMPLED (L) | MEASURED COUNTS (#) | CONCENTRATION (#/M ³) | 95% UCL (#/M ³) | ALLOWABLE RANGE (#/M ³) | PASS/FAIL |
|-------------------|-------------|---------------------|-----------------------------------|-----------------------------|-------------------------------------|-----------|
| 30 | 85 | 0 | 0.00 | 35.3 | ≤70.7 | Pass |

SAMPLING FLOW RATE (L/MIN)

| NOMINAL | ACTUAL | ERROR | ALLOWABLE RANGE | PASS/FAIL |
|---------|--------|-------|-----------------|-----------|
| 2.83 | 2.83 | 0.0 % | ± 5% | Pass |

SAMPLING TIME †

| MEASURED | ALLOWABLE RANGE | PASS/FAIL |
|----------|-----------------|-----------|
| < ± 0.1% | ± 1% | Pass |

RESPONSE RATE †

| MEASURED | ALLOWABLE RANGE | PASS/FAIL |
|----------|-----------------|-----------|
| 0.08% | ≤ 0.5% | Pass |

MAXIMUM PARTICLE CONCENTRATION †

210000000 #/m³ @10% Coincidence Loss

† Tested and verified during product development

CALIBRATION INTERVAL

| CALIBRATION DATE | EXPIRATION DATE |
|------------------|-----------------|
| October 4, 2017 | October 4, 2018 |

Model 9306-V2 SN 93061247001 Wednesday, October 04, 2017 6:16:33 PM

Page 2 of 2



CERTIFICATE OF CALIBRATION

TSI Incorporated, 500 Cardigan Road, Shoreview, MN 55126 USA
Tel: 1-800-874-2811 1-651-490-2811 Fax: 1-651-490-3824 http://www.tsi.com

| ENVIRONMENT CONDITION | | |
|-----------------------|---------------|------------|
| TEMPERATURE | 74.6 (23.7) | °F (°C) |
| RELATIVE HUMIDITY | 31 | %RH |
| BAROMETRIC PRESSURE | 29.26 (990.9) | inHg (hPa) |

| | |
|------------------|--------------------|
| MODEL | 9306-V2 |
| SERIAL NUMBER | 93061247001 |
| CUSTOMER INST ID | |

| | |
|---|--|
| <input checked="" type="checkbox"/> AS LEFT | <input checked="" type="checkbox"/> IN TOLERANCE |
| <input type="checkbox"/> AS FOUND | <input type="checkbox"/> OUT OF TOLERANCE |

| AEROTRAK CALIBRATION KIT | | | |
|--------------------------|-----------|----------------------|----------------------|
| MEASUREMENT VARIABLE | SYSTEM ID | DATE LAST CALIBRATED | CALIBRATION DUE DATE |
| FLOW METER | E003739 | 6/30/2017 | 12/31/2017 |
| 7201-02F | E005520 | 9/18/2017 | 3/31/2018 |
| FLOW METER | E005633 | 7/12/2017 | 1/31/2018 |

| PARTICLE STANDARDS | | | | |
|--------------------|----------------------|--------------------|---------|-----------------|
| PARTICLE SIZE | STANDARD UNCERTAINTY | STANDARD DEVIATION | LOT NO. | EXPIRATION DATE |
| 0.303 µm | 0.003 µm | 0.0047 µm | 174664 | 10/31/2019 |
| 0.508 µm | 0.004 µm | 0.0085 µm | 168223 | 4/30/2019 |
| 0.994 µm | 0.0075 µm | 0.010 µm | 171667 | 7/31/2019 |
| 3.000 µm | 0.01 µm | 0.03 µm | 167683 | 3/31/2019 |
| 5.020 µm | 0.015 µm | 0.06 µm | 179268 | 1/31/2020 |
| 9.850 µm | 0.04 µm | 0.13 µm | 172685 | 8/31/2019 |

TSI does hereby certify that the calibration performed on the above described instrument meets the requirements of ISO 21501-4. TSI does hereby certify that the above described instrument conforms to the original manufacturer's specification (not applicable to As Found data) and has been calibrated using standards whose accuracies are traceable to the United States National Institute of Standards and Technology (NIST) or has been verified with respect to instrumentation whose accuracy is traceable to NIST, or is derived from accepted values of physical constants. TSI is registered to ISO-9001:2015.

Charles Truano
CALIBRATED

October 4, 2017

DATE

Instrument TSI AeroTrak 9306-V2 Particle Counter
 Date 10/25/2017
 Start Time 11:40
 End Time 14:22

| | Minimum | Average | Maximum | Std. Dev. | Std. Err. |
|-----------|---------|----------|---------|-----------|-----------|
| Channel 1 | 0 | 14843.09 | 315548 | 53990.64 | 9259.32 |
| Channel 2 | 0 | 7458.91 | 144170 | 24512.33 | 4203.83 |
| Channel 3 | 0 | 5186 | 105477 | 17927.47 | 3074.54 |
| Channel 4 | 0 | 2390.35 | 48410 | 8268.6 | 1418.05 |
| Channel 5 | 0 | 1584.88 | 33216 | 5681.39 | 974.35 |
| Channel 6 | 0 | 732.68 | 13781 | 2376.45 | 407.56 |

| Record # | Location | Instrument Status | Sample Time | Laser Status | Cumulative | | Flow Status | Volume(m3) |
|----------|----------------|-------------------|-------------|--------------|------------|--------|-------------|------------|
| | | | | | 0.3µm | 0.5µm | | |
| 1 | Purge | (ok) | 0:05:00 | OK | 71 | 71 | OK | 0.01 |
| 2 | Zero | (ok) | 0:05:00 | OK | 0 | 0 | OK | 0.01 |
| 4 | Near Window | (ok) | 0:02:00 | OK | 6184 | 4770 | OK | 0.01 |
| 5 | Near Window | (ok) | 0:02:00 | OK | 12014 | 9187 | OK | 0.01 |
| 6 | Near Window | (ok) | 0:02:00 | OK | 13958 | 12367 | OK | 0.01 |
| 8 | Near Window | (ok) | 0:02:00 | OK | 1943 | 1590 | OK | 0.01 |
| | Avg | | | | 8525 | 5492 | 6979 | 4755 |
| 9 | Near Fume Hood | (ok) | 0:02:00 | OK | 707 | 177 | OK | 0.01 |
| 10 | Near Fume Hood | (ok) | 0:02:00 | OK | 2297 | 1590 | OK | 0.01 |
| 11 | Near Fume Hood | (ok) | 0:02:00 | OK | 6184 | 5654 | OK | 0.01 |
| 12 | Near Fume Hood | (ok) | 0:02:00 | OK | 10601 | 9541 | OK | 0.01 |
| 13 | Near Fume Hood | (ok) | 0:02:00 | OK | 3887 | 3004 | OK | 0.01 |
| | Avg | | | | 4735 | 3853 | 3993 | 3703 |
| 14 | Center of Lab | (ok) | 0:02:00 | OK | 530 | 0 | OK | 0.01 |
| 15 | Center of Lab | (ok) | 0:02:00 | OK | 707 | 707 | OK | 0.01 |
| 16 | Center of Lab | (ok) | 0:02:00 | OK | 1237 | 707 | OK | 0.01 |
| 17 | Center of Lab | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| 18 | Center of Lab | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| | Avg | | | | 495 | 521 | 283 | 387 |
| 24 | In Flow Bench | (ok) | 0:02:00 | OK | 883 | 353 | OK | 0.01 |
| 25 | In Flow Bench | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| 26 | In Flow Bench | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| 27 | In Flow Bench | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| 28 | In Flow Bench | (ok) | 0:02:00 | OK | 0 | 0 | OK | 0.01 |
| | Avg | | | | 177 | 395 | 71 | 158 |
| 29 | Near Entrance | (ok) | 0:02:00 | OK | 45053 | 14488 | OK | 0.01 |
| 30 | Near Entrance | (ok) | 0:02:00 | OK | 32509 | 9541 | OK | 0.01 |
| 31 | Near Entrance | (ok) | 0:02:00 | OK | 13958 | 6890 | OK | 0.01 |
| 32 | Near Entrance | (ok) | 0:02:00 | OK | 2827 | 2473 | OK | 0.01 |
| 33 | Near Entrance | (ok) | 0:02:00 | OK | 2473 | 1413 | OK | 0.01 |
| | Avg | | | | 19364 | 18846 | 6961 | 5344 |
| 34 | Changing Zone | (ok) | 0:02:00 | OK | 315548 | 144170 | OK | 0.01 |